



Accreditation Scope

Autocontrol Process Instrumentation FZC, NAL 244
Calibration Laboratory, (ISO/IEC 17025:2017)

Plot No-1H-08B, Hamriyah Free Zone-1, Sharjah, Sharjah, United Arab Emirates

Issue Date: 15/01/2026

Expiry Date: 14/01/2029

Issue No: 1

Calibration Field/ Quantity/ Property	Measurand / Equipment	Measurement Range	Expanded Measurement Uncertainty (k=2)	Calibration Method (Standard/ Internal Procedure)	Permanent lab (P)/Client-site
Pressure	Pneumatic Pressure Gauge/ Differential Pressure Gauge	(50 to 500) mbar	0.004 bar	API-QLT-WI-01 (based on DKD-R 6-1)	P
Pressure	Pneumatic Pressure Gauge/ Differential Pressure Gauge	(100 to 1000) mbar	0.004 bar	API-QLT-WI-01 (based on DKD-R 6-1)	P
Pressure	Pneumatic Pressure Gauge/ Differential Pressure Gauge	(>1 to 20) bar	0.008 bar	API-QLT-WI-01 (based on DKD-R 6-1)	P
Pressure	Pneumatic Pressure Gauge/ Differential Pressure Gauge	(20 to 210) bar	0.03 bar	API-QLT-WI-01 (based on DKD-R 6-1)	P

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.