



Accreditation Scope

Autocontrol Process Instrumentation FZC, NAL 244
Calibration Laboratory, (ISO/IEC 17025:2017)

Plot No-1H-08B, Hamriyah Free Zone-1, Sharjah, Sharjah, United Arab Emirates

Issue Date: 15/01/2026

Expiry Date: 14/01/2029

Issue No: 1

Pressure	Pneumatic Pressure Transmitter / Differential Pressure Transmitter with Indicator	(-0.9 to -0.1) bar	0.004 bar	API-QLT-WI-02 (based on DKD-R 6-1)	P
Pressure	Pneumatic Pressure Transmitter / Differential Pressure Transmitter with Indicator	(0 to 1000) mbar	0.004 bar	API-QLT-WI-02 (based on DKD-R 6-1)	P
Pressure	Pneumatic Pressure Transmitter / Differential Pressure Transmitter with Indicator	(>1 to 20) bar	0.009 bar	API-QLT-WI-02 (based on DKD-R 6-1)	P
Pressure	Pneumatic Pressure Transmitter / Differential Pressure Transmitter with Indicator	(20 to 210) bar	0.03 bar	API-QLT-WI-02 (based on DKD-R 6-1)	P

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.