



Accreditation Scope

Autocontrol Process Instrumentation FZC, NAL 244
Calibration Laboratory, (ISO/IEC 17025:2017)

Plot No-1H-08B, Hamriyah Free Zone-1, Sharjah, Sharjah, United Arab Emirates

Issue Date: 15/01/2026

Expiry Date: 14/01/2029

Issue No: 1

Pressure	Hydraulic Pressure Transmitter / Differential Pressure Transmitter with Indicator	(35 to 100) bar	0.04 bar	API-QLT-WI-02 (based on DKD-R 6-1)	P
Pressure	Hydraulic Pressure Transmitter / Differential Pressure Transmitter with Indicator	(100 to 700) bar	0.2 bar	API-QLT-WI-02 (based on DKD-R 6-1)	P
Pressure	Vacuum Gauge (Analog & Digital)	(0 to -0.9) bar	0.004 bar	API-QLT-WI-01 (based on DKD-R 6-1)	P
Thermodynamics and temperature measurements	Temperature Indicator with RTD	(-40 to 100) °C	0.11 °C	API-QLT-WI-03 (based on DKD-R 5-1)	P
Thermodynamics and temperature measurements	Temperature Indicator with RTD	(>100 to 400) °C	0.21 °C	API-QLT-WI-03 (based on DKD-R 5-1)	P
Thermodynamics and temperature measurements	Temperature Indicator with RTD	(>400 to 650) °C	0.63 °C	API-QLT-WI-03 (based on DKD-R 5-1)	P

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.